

お客様各位

2025 年 4 月 25 日

株式会社島津製作所 分析計測事業部  
ダイアグノスティクス事業統括部  
臨床・感染症検査ビジネスユニット

### MultiNA 用試薬キットの原料変更のお知らせ

平素より弊社製品 MultiNA をご愛顧いただき誠にありがとうございます。

MultiNA 用試薬キットの原材料が一部製造中止となったため、代替品の使用とそれに伴い一部組成を変更しておりますことをお知らせします。当社管理基準における従来品との同等性を確認しておりますが、個々の分析においてピーク形状等の結果に影響を与える可能性があります。保存性に対する影響はありません。また、部品番号の変更はありません。引き続き、ご愛顧のほどよろしくお願い申し上げます。

#### 原材料が変更となる試薬キット

No.	部品番号	品名	ロット番号
1	292-27910-91	DNA-500 キット	Lot#2081 (2024 年 8 月生産分)以降のマーカ溶液
2	292-27911-91	DNA-1000 キット	Lot#2031 (2025 年 3 月生産分)以降の分離バッファとマーカ溶液
3	292-27912-91	DNA-2500 キット	Lot#3051 (2025 年 5 月生産分)以降の分離バッファとマーカ溶液
4	292-36600-91	DNA-12000 キット	Lot#2111 (2024 年 11 月生産分)以降の分離バッファ
5	292-27913-91	RNA キット	Lot#未定 (2025 年 10 月生産分)以降の分離バッファを予定

Dear valued customer,

April 25, 2025

Subject: Notice of change in ingredients for MultiNA Reagent Kit

Thank you for your continued patronage of our product MultiNA.

We would like to inform you that some of the raw materials for the reagent kit for MultiNA have been discontinued, so we are using alternative materials and have changed some of the composition accordingly. We have confirmed that it meets our quality standards and is equivalent to conventional products, however, it may affect individual analytical results such as peak shape. This will not affect storage stability. In addition, there will be no changes to part numbers.

We appreciate your continued support.

Reagent kits with changed raw materials

No.	Part Number	Product Name	Lot Number
1	292-27910-91	DNA-500 Kit	Lot#2081 (manufactured in August 2024) or later, Marker solution
2	292-27911-91	DNA-1000 Kit	Lot#2031 (manufactured in March 2025) or later, Buffer and Marker solution
3	292-27912-91	DNA-2500 Kit	Lot#3051 (manufactured in May 2025) or later, Buffer and Marker solution
4	292-36600-91	DNA-12000 Kit	Lot#2111 (manufactured in November 2024) or later, Separation Buffer
5	292-27913-91	RNA Kit	TBD (scheduled for production from October 2025), Separation Buffer